

How to use LR-HSQMBC: Observation of very small couplings

Product used : Nuclear Magnetic Resonance (NMR)

The HMBC (Heteronuclear Multiple Bond Correlation) experiment is the most widely used technique for observing heteronuclear longrange correlations. HMBC mainly observes 2- and 3-bond correlations; however, 4-bond correlations can sometimes be observed as well. The correlations observed in an HMBC spectrum strongly depend on the matching of the transfer delay used in the experiment with the long-range heteronuclear *J*-couplings that exist in the molecule.

HMBC correlation signal is anti-phase as shown in Fig. 1a), however, this is not usually observed in HMBC spectra, because HMBC data are usually processed in absolute-value mode. However, this anti-phase nature of HMBC signal may lead to partial cancellation and signal reduction for very small long-range couplings, as shown in Fig. 1b).

To overcome this problem, the D-HMBC (Decoupled HMBC)^{$\times1$} and LR-HSQMBC (Long-Range Heteronuclear Single Quantum Multiple Bond Correlation) experiments were developed. In D-HMBC and LR-HSQMBC, signals appear as in-phase doublets in the direct (F_2) dimension, as shown in Fig. 2a. Therefore, there is no signal cancellation in the case of very small couplings, and heteronuclear decoupling can also be applied in order to increase sensitivity. For these reasons, the experiments offer higher sensitivity than the traditional HMBC experiment.

On the other hand, the D-HMBC experiment suffers from signal splitting by J_{HH} in the indirect dimension, which reduces the resolution of the spectrum. Here we will introduce the LR-HSQMBC experiment in detail because of its superior resolution.



Fig. 3 shows the pulse sequence diagram of LR-HSQMBC experiment. To observe heteronuclear long-range correlations that are difficult to observe with HMBC, ${}^{n}J_{XH}$ should be set to a small value, such as 2 Hz. It is a standard experiment from NMR software Delta V.5.2.0. The name of the pulse program is **Irhsqmbc.jxp**.



Fig. 3: Pulse sequence of LR-HSQMBC

Reference of LR-HSQMBC

R. Thomas Williamson, Alexei V. Buevich, Gary E. Martin, Teodor Parella, J. Org. Chem. 2014, 79, 3887-3894. R. Thomas Williamson, Alexei V. Buevich, Gary E. Martin, Tetrahedron Letters 2014, 55, 3365-3366.



Here we compare the aromatic region of ¹H-¹³C HMBC and ¹H-¹³C LR-HSQMBC spectra recorded on a sample of vinclozolin. In general, ³J_{CH} couplings are large on aromatic rings, and hence correlation signals arising from 3-bond couplings are of high intensity.

¹H-¹³C HMBC

Fig. 4 shows two results obtained with the HMBC experiment. Although, the long-range delay was optimised to 8 Hz (a) and 2 Hz (b), respectively, the structural information obtained from the spectra is the same. In both cases, all 2-bond and 3-bond correlations of protons H4 and H5 were observed.



Fig. 4: Expansions of HMBC: a) 8 scans, long_range_j: 8 Hz , b) 32 scans, long_range_j: 2 Hz

¹H-¹³C LR-HSQMBC

Fig. 5 shows two results obtained with the LR-HSQMBC experiment. The spectrum in a) shows the same result as HMBC in Fig. 4a). However, the H5/C6 correlation in the blue oval arising from ${}^{4}J_{CH}$ was not observed with HMBC in Fig. 4b). The coupling constant between H5 and C6 is 2.2 Hz. While HMBC is an absolute-value experiment, LR-HSQMBC is a phase-sensitive experiment, and hence the number of scans in LR-HSQMBC was set to half of scans in HMBC in order to set the same experimental time in both experiments.





Sample : 5mg Vinclozolin in DMSO-d6 Instrument : JNM-ECZ500R, ROYALPROBE™ HFX



Since ¹H-¹⁵N long-range couplings are similar in magnitude to ¹H-¹³C long-range couplings, the LR-HSQMBC experiment also has potential to improve the detection of ¹H-¹⁵N long-range correlations. As ³J_{NH} couplings have a larger variation than ³J_{CH}, the entire range of ³J_{NH} couplings cannot be completely covered with a single HMBC experiment. It is worth mentioning that structure elucidation of compounds containing nitrogen strongly relies on 2D ¹H-¹⁵N correlation experiments, because the direct detection of ¹⁵N is very challenging at natural abundance.

¹H-¹⁵N HMBC and ¹H-¹⁵N LR-HSQMBC

Fig. 6 shows two results obtained with the HMBC and LR-HSQMBC experiments. The only H4/N correlation can be observed in the ¹H-¹⁵N HMBC spectrum a). On the other hand, H4/N and H1/N (${}^{4}J_{NH}$) can be observed in the LR-HSQMBC spectrum b). The size of H4/N coupling is 4.2 Hz and the size of H1/N coupling is 1.5 Hz.



Fig. 6: a) ¹H-¹⁵N HMBC, 32 scans, long_range_i: 2 Hz , b) ¹H-¹⁵N LR-HSQMBC, 16 scans, long_range_i: 2 Hz

Sample : 5mg Vinclozolin in DMSO-d6 Instrument : JNM-ECZ500R, ROYALPROBE[™] HFX

Features and main application of each experiment

The principal features of heteronuclear long-range correlation experiments can be summarised as follows:

At first, HMBC correlations should be used for analysis. If you cannot differentiate between ${}^{2}J_{CH}$ and ${}^{3}J_{CH}$ HMBC correlations, consider H2BC or 1,1-ADEQUATE. H2BC is suitable for proton-rich samples, while 1,1-ADEQUATE can also be used for proton-deficient samples.

If the structural information obtained from HMBC is inadequate for analysis due to many quaternary carbons and hetero atoms in the molecule, then consider LR-HSQMBC.

HMBC --- Observe mainly ${}^{2}J_{CH}$ and ${}^{3}J_{CH}$ correlations.

H2BC $^{\otimes 2}$ --- Selectively observe ${}^{2}J_{CH}$ correlations. Correlations of quaternary carbon atoms cannot be observed.

1,1-ADEQUATE ³ --- Observe ¹J_{CH} and ²J_{CH} correlations. Correlations of quaternary carbons can be observed. Very low sensitivity.

LR-HSQMBC --- Observe correlations close to the set *J*-value. However, it is difficult to determine the number of chemical bonds between the interacting partners.

Reference

JEUI

%1 Kazui Furihata, Haruo Seto, Tetrahedron Letters 1995, 36, 2817-2820.
%2 JEOL Application note NM140007.

X3 JEOL Application note NM160002.

JEOL Ltd.

Copyright © 2021 JEOL Ltd.

Certain products in this brochure are controlled under the "Foreign Exchange and Foreign Trade Law" of Japan in compliance with international security export control. JEOL Ltd. must provide the Japanese Government with "End-user's Statement of Assurance" and "End-use Certificate" in order to obtain the export license needed for export from Japan. If the product to be exported is in this category, the end user will be asked to fill in these certificate forms.

3-1-2 Musashino Akishima Tokyo 196-8558 Japan Sales Division Tel. +81-3-6262-3560 Fax. +81-3-6262-3577 www.jeol.com ISO 9001 • ISO 14001 Certified





• KOREA • MALAYSIA • MEXICO • RUSSIA • SCANDINAVI